



Exova Jones Environmental

Registered Office: Exova Environmental UK Limited, 10 Lower Grosvenor Place, London, SW1W 0EN. Reg No. 11371415

Unit 3 Deeside Point
Zone 3
Deeside Industrial Park
Deeside
CH5 2UA

E-Geo Solutions
Oak House
Groes Lwyd
Abergele
LL22 7SU

Tel: +44 (0) 1244 833780

Fax: +44 (0) 1244 833781



Attention :	Huw Littler-Jones
Date :	26th November, 2018
Your reference :	Coed Poeth
Our reference :	Test Report 18/18460 Batch 1
Location :	Tabor Coed Poeth
Date samples received :	15th November, 2018
Status :	Final report
Issue :	1

One sample was received for analysis on 15th November, 2018 of which one was scheduled for analysis. Please find attached our Test Report which should be read with notes at the end of the report and should include all sections if reproduced. Interpretations and opinions are outside the scope of any accreditation, and all results relate only to samples supplied.

All analysis is carried out on as received samples and reported on a dry weight basis unless stated otherwise. Results are not surrogate corrected.

Compiled By:

Phil Sommerton BSc

Project Manager

Client Name: E-Geo Solutions
Reference: Coed Poeth
Location: Tabor Coed Poeth
Contact: Huw Littler-Jones
JE Job No.: 18/18460

Report : Solid

Solids: V=60g VOC jar, J=250g glass jar, T=plastic tub

J E Sample No.	1										Please see attached notes for all abbreviations and acronyms		
Sample ID	WAC1												
Depth	0.30												
COC No / misc													
Containers	B												
Sample Date	14/11/2018												
Sample Type	Clay												
Batch Number	1												
Date of Receipt	15/11/2018										LOD/LOR	Units	Method No.
PAH MS													
Naphthalene ^{#M}	<0.04										<0.04	mg/kg	TM4/PM8
Acenaphthylene	<0.03										<0.03	mg/kg	TM4/PM8
Acenaphthene ^{#M}	<0.05										<0.05	mg/kg	TM4/PM8
Fluorene ^{#M}	<0.04										<0.04	mg/kg	TM4/PM8
Phenanthrene ^{#M}	0.14										<0.03	mg/kg	TM4/PM8
Anthracene [#]	0.05										<0.04	mg/kg	TM4/PM8
Fluoranthene ^{#M}	0.71										<0.03	mg/kg	TM4/PM8
Pyrene [#]	0.64										<0.03	mg/kg	TM4/PM8
Benzo(a)anthracene [#]	0.40										<0.06	mg/kg	TM4/PM8
Chrysene ^{#M}	0.32										<0.02	mg/kg	TM4/PM8
Benzo(bk)fluoranthene ^{#M}	0.59										<0.07	mg/kg	TM4/PM8
Benzo(a)pyrene [#]	0.33										<0.04	mg/kg	TM4/PM8
Indeno(123cd)pyrene ^{#M}	0.21										<0.04	mg/kg	TM4/PM8
Dibenzo(ah)anthracene [#]	0.05										<0.04	mg/kg	TM4/PM8
Benzo(ghi)perylene [#]	0.20										<0.04	mg/kg	TM4/PM8
Coronene	<0.04										<0.04	mg/kg	TM4/PM8
PAH 17 Total	3.64										<0.64	mg/kg	TM4/PM8
Benzo(b)fluoranthene	0.42										<0.05	mg/kg	TM4/PM8
Benzo(k)fluoranthene	0.17										<0.02	mg/kg	TM4/PM8
PAH Surrogate % Recovery	102										<0	%	TM4/PM8
Mineral Oil (C10-C40)	<30										<30	mg/kg	TM5/PM8/PM16
MTBE [#]	<5										<5	ug/kg	TM31/PM12
Benzene [#]	<5										<5	ug/kg	TM31/PM12
Toluene [#]	<5										<5	ug/kg	TM31/PM12
Ethylbenzene [#]	<5										<5	ug/kg	TM31/PM12
m/p-Xylene [#]	<5										<5	ug/kg	TM31/PM12
o-Xylene [#]	<5										<5	ug/kg	TM31/PM12
PCB 28 [#]	<5										<5	ug/kg	TM17/PM8
PCB 52 [#]	<5										<5	ug/kg	TM17/PM8
PCB 101 [#]	<5										<5	ug/kg	TM17/PM8
PCB 118 [#]	<5										<5	ug/kg	TM17/PM8
PCB 138 [#]	<5										<5	ug/kg	TM17/PM8
PCB 153 [#]	<5										<5	ug/kg	TM17/PM8
PCB 180 [#]	<5										<5	ug/kg	TM17/PM8
Total 7 PCBs [#]	<35										<35	ug/kg	TM17/PM8
Natural Moisture Content	18.8										<0.1	%	PM4/PM0
Total Organic Carbon [#]	1.35										<0.02	%	TM21/PM24

Mass of sample taken (kg)	-	Moisture Content Ratio (%) =	18.8		
Mass of dry sample (kg) =	0.09	Dry Matter Content Ratio (%) =	84.1		
Particle Size <4mm =	>95%				
JEFL Job No	18/18460		Landfill Waste Acceptance Criteria Limits		
Sample No	1				
Client Sample No	WAC1		Inert Waste Landfill	Stable Non-reactive Hazardous Waste in Non-Hazardous Landfill	Hazardous Waste Landfill
Depth/Other	0.30				
Sample Date	14/11/2018				
Batch No	1				
Solid Waste Analysis					
Total Organic Carbon (%)	1.35				
Loss on Ignition (%)	4.8		-	-	10
Sum of BTEX (mg/kg)	<0.025		6	-	-
Sum of 7 PCBs (mg/kg)	<0.035		1	-	-
Mineral Oil (mg/kg)	<30		500	-	-
PAH Sum of 17(mg/kg)	3.64		100	-	-
pH (pH Units)	8.49		-	>6	-
ANC to pH 7 (mol/kg)	NDP		-	to be evaluated	to be evaluated
ANC to pH 4 (mol/kg)	0.05		-	to be evaluated	to be evaluated
Eluate Analysis	10:1 conc ⁿ leached		Limit values for compliance leaching test using BS EN 12457-2 at L/S 10 l/kg		
	C ₁₀	A ₁₀			
	mg/l	mg/kg			
Arsenic	<0.0025	<0.025	0.5	2	25
Barium	<0.003	<0.03	20	100	300
Cadmium	<0.0005	<0.005	0.04	1	5
Chromium	<0.0015	<0.015	0.5	10	70
Copper	<0.007	<0.07	2	50	100
Mercury	<0.001	<0.01	0.01	0.2	2
Molybdenum	0.011	0.11	0.5	10	30
Nickel	<0.002	<0.02	0.4	10	40
Lead	<0.005	<0.05	0.5	10	50
Antimony	<0.002	<0.02	0.06	0.7	5
Selenium	<0.003	<0.03	0.1	0.5	7
Zinc	<0.003	<0.03	4	50	200
Chloride	<0.3	<3	800	15000	25000
Fluoride	<0.3	<3	10	150	500
Sulphate as SO4	0.73	7.3	1000	20000	50000
Total Dissolved Solids	57	570	4000	60000	100000
Phenol	<0.01	<0.1	1	-	-
Dissolved Organic Carbon	3	30	500	800	1000

Matrix : Solid

[illegible]

Client Name: E-Geo Solutions
Reference: Coed Poeth
Location: Tabor Coed Poeth
Contact: Huw Littler-Jones

[illegible]

Please note that only samples that are deviating are mentioned in this report. If no samples are listed it is because none were deviating. Only analyses which are accredited are recorded as deviating if set criteria are not met.

NOTES TO ACCOMPANY ALL SCHEDULES AND REPORTS

JE Job No.: 18/18460

SOILS

Please note we are only MCERTS accredited (UK soils only) for sand, loam and clay and any other matrix is outside our scope of accreditation.

Where an MCERTS report has been requested, you will be notified within 48 hours of any samples that have been identified as being outside our MCERTS scope. As validation has been performed on clay, sand and loam, only samples that are predominantly these matrices, or combinations of them will be within our MCERTS scope. If samples are not one of a combination of the above matrices they will not be marked as MCERTS accredited.

It is assumed that you have taken representative samples on site and require analysis on a representative subsample. Stones will generally be included unless we are requested to remove them.

All samples will be discarded one month after the date of reporting, unless we are instructed to the contrary.

If you have not already done so, please send us a purchase order if this is required by your company.

Where appropriate please make sure that our detection limits are suitable for your needs, if they are not, please notify us immediately.

All analysis is reported on a dry weight basis unless stated otherwise. Results are not surrogate corrected. Samples are dried at 35°C ±5°C unless otherwise stated. Moisture content for CEN Leachate tests are dried at 105°C ±5°C.

Where Mineral Oil or Fats, Oils and Grease is quoted, this refers to Total Aliphatics C10-C40.

Where a CEN 10:1 ZERO Headspace VOC test has been carried out, a 10:1 ratio of water to wet (as received) soil has been used.

% Asbestos in Asbestos Containing Materials (ACMs) is determined by reference to HSG 264 The Survey Guide - Appendix 2 : ACMs in buildings listed in order of ease of fibre release.

Negative Neutralization Potential (NP) values are obtained when the volume of NaOH (0.1N) titrated (pH 8.3) is greater than the volume of HCl (1N) to reduce the pH of the sample to 2.0 - 2.5. Any negative NP values are corrected to 0.

The calculation of Pyrite content assumes that all oxidisable sulphides present in the sample are pyrite. This may not be the case. The calculation may be an overestimate when other sulphides such as Barite (Barium Sulphate) are present.

WATERS

Please note we are not a UK Drinking Water Inspectorate (DWI) Approved Laboratory .

ISO17025 accreditation applies to surface water and groundwater and usually one other matrix which is analysis specific, any other liquids are outside our scope of accreditation.

As surface waters require different sample preparation to groundwaters the laboratory must be informed of the water type when submitting samples.

Where Mineral Oil or Fats, Oils and Grease is quoted, this refers to Total Aliphatics C10-C40.

DEVIATING SAMPLES

Samples must be received in a condition appropriate to the requested analyses. All samples should be submitted to the laboratory in suitable containers with sufficient ice packs to sustain an appropriate temperature for the requested analysis. If this is not the case you will be informed and any test results that may be compromised highlighted on your deviating samples report.

SURROGATES

Surrogate compounds are added during the preparation process to monitor recovery of analytes. However low recovery in soils is often due to peat, clay or other organic rich matrices. For waters this can be due to oxidants, surfactants, organic rich sediments or remediation fluids. Acceptable limits for most organic methods are 70 - 130% and for VOCs are 50 - 150%. When surrogate recoveries are outside the performance criteria but the associated AQC passes this is assumed to be due to matrix effect. Results are not surrogate corrected.

DILUTIONS

A dilution suffix indicates a dilution has been performed and the reported result takes this into account. No further calculation is required.

BLANKS

Where analytes have been found in the blank, the sample will be treated in accordance with our laboratory procedure for dealing with contaminated blanks.

NOTE

Data is only reported if the laboratory is confident that the data is a true reflection of the samples analysed. Data is only reported as accredited when all the requirements of our Quality System have been met. In certain circumstances where all the requirements of the Quality System have not been met, for instance if the associated AQC has failed, the reason is fully investigated and documented. The sample data is then evaluated alongside the other quality control checks performed during analysis to determine its suitability. Following this evaluation, provided the sample results have not been effected, the data is reported but accreditation is removed. It is a UKAS requirement for data not reported as accredited to be considered indicative only, but this does not mean the data is not valid.

Where possible, and if requested, samples will be re-extracted and a revised report issued with accredited results. Please do not hesitate to contact the laboratory if further details are required of the circumstances which have led to the removal of accreditation.

REPORTS FROM THE SOUTH AFRICA LABORATORY

Any method number not prefixed with SA has been undertaken in our UK laboratory unless reported as subcontracted.

Please include all sections of this report if it is reproduced

ABBREVIATIONS and ACRONYMS USED

#	ISO17025 (UKAS Ref No. 4225) accredited - UK.
SA	ISO17025 (SANAS Ref No.T0729) accredited - South Africa.
B	Indicates analyte found in associated method blank.
DR	Dilution required.
M	MCERTS accredited.
NA	Not applicable
NAD	No Asbestos Detected.
ND	None Detected (usually refers to VOC and/SVOC TICs).
NDP	No Determination Possible
SS	Calibrated against a single substance
SV	Surrogate recovery outside performance criteria. This may be due to a matrix effect.
W	Results expressed on as received basis.
+	AQC failure, accreditation has been removed from this result, if appropriate, see 'Note' on previous page.
++	Result outside calibration range, results should be considered as indicative only and are not accredited.
*	Analysis subcontracted to an Exova Jones Environmental approved laboratory.
AD	Samples are dried at 35°C ±5°C
CO	Suspected carry over
LOD/LOR	Limit of Detection (Limit of Reporting) in line with ISO 17025 and MCERTS
ME	Matrix Effect
NFD	No Fibres Detected
BS	AQC Sample
LB	Blank Sample
N	Client Sample
TB	Trip Blank Sample
OC	Outside Calibration Range

JE Job No: 18/18460

Test Method No.	Description	Prep Method No. (if appropriate)	Description	ISO 17025 (UKAS/ANAS)	MCERTS (UK soils only)	Analysis done on As Received (AR) or Dried (AD)	Reported on dry weight basis
PM4	Gravimetric measurement of Natural Moisture Content and % Moisture Content at either 35°C or 105°C. Calculation based on ISO 11465 and BS1377.	PM0	No preparation is required.			AR	
TM4	Modified USEPA 8270 method for the solvent extraction and determination of 16 PAHs by GC-MS.	PM8	End over end extraction of solid samples for organic analysis. The solvent mix varies depending on analysis required.			AR	Yes
TM4	Modified USEPA 8270 method for the solvent extraction and determination of 16 PAHs by GC-MS.	PM8	End over end extraction of solid samples for organic analysis. The solvent mix varies depending on analysis required.	Yes		AR	Yes
TM4	Modified USEPA 8270 method for the solvent extraction and determination of 16 PAHs by GC-MS.	PM8	End over end extraction of solid samples for organic analysis. The solvent mix varies depending on analysis required.	Yes	Yes	AR	Yes
TM5	Modified USEPA 8015B method for the determination of solvent Extractable Petroleum Hydrocarbons (EPH) with carbon banding within the range C8-C40 GC-FID.	PM8/PM16	End over end extraction of solid samples for organic analysis. The solvent mix varies depending on analysis required/Fractionation into aliphatic and aromatic fractions using a Rapid Trace SPE.			AR	Yes
PM13	A visual examination of the solid sample is carried out to ascertain sample make up, colour and any other inclusions. This is not a geotechnical description.	PM0	No preparation is required.			AR	
TM17	Modified US EPA method 8270. Determination of specific Polychlorinated Biphenyl congeners by GC-MS.	PM8	End over end extraction of solid samples for organic analysis. The solvent mix varies depending on analysis required.	Yes		AR	Yes
TM20	Modified BS 1377-3: 1990/USEPA 160.3 Gravimetric determination of Total Dissolved Solids/Total Solids	PM0	No preparation is required.			AR	Yes
TM21	Modified BS 7755-3:1995, ISO10694:1995 Determination of Total Organic Carbon or Total Carbon by combustion in an Eltra TOC furnace/analyser in the presence of oxygen. The CO ₂ generated is quantified using infra-red detection. Organic Matter (SOM) calculated as per EA MCERTS Chemical Testing of Soil, March 2012 v4.	PM24	Dried and ground solid samples are washed with hydrochloric acid, then rinsed with deionised water to remove the mineral carbon before TOC analysis.	Yes		AD	Yes
TM22	Modified USEPA 160.4. Gravimetric determination of Loss on Ignition by temperature controlled Muffle Furnace (450°C)	PM0	No preparation is required.	Yes		AD	Yes

JE Job No: 18/18460

Test Method No.	Description	Prep Method No. (if appropriate)	Description	ISO 17025 (UKAS/ANAS)	MCERTS (UK soils only)	Analysis done on As Received (AR) or Dried (AD)	Reported on dry weight basis
TM26	Determination of phenols by Reversed Phased High Performance Liquid Chromatography and Electro-Chemical Detection.	PM0	No preparation is required.			AR	Yes
TM30	Determination of Trace Metal elements by ICP-OES (Inductively Coupled Plasma - Optical Emission Spectrometry). Modified US EPA Method 200.7, 6010B and BS EN ISO 11885 2009	PM17	Modified method EN12457-2 As received solid samples are leached with water in a 10:1 water to soil ratio for 24 hours, the moisture content of the sample is included in the ratio.	Yes		AR	Yes
TM31	Modified USEPA 8015B. Determination of Methyltertbutylether, Benzene, Toluene, Ethylbenzene and Xylene by headspace GC-FID.	PM12	Modified US EPA method 5021. Preparation of solid and liquid samples for GC headspace analysis.			AR	Yes
TM31	Modified USEPA 8015B. Determination of Methyltertbutylether, Benzene, Toluene, Ethylbenzene and Xylene by headspace GC-FID.	PM12	Modified US EPA method 5021. Preparation of solid and liquid samples for GC headspace analysis.	Yes		AR	Yes
TM38	Soluble Ion analysis using Discrete Analyser. Modified US EPA methods 325.2 (Chloride), 375.4 (Sulphate), 365.2 (o-Phosphate), 353.1 (TON), 354.1 (Nitrite), 350.1 (NH4+) comparable to BS ISO 15923-1, 7196A (Hex Cr)	PM0	No preparation is required.	Yes		AR	Yes
TM60	TC/TOC analysis of Waters by High Temperature Combustion followed by NDIR detection. Based on the following modified standard methods: USEPA 9060, APHA Standard Methods for Examination of Water and Wastewater 5310B, ASTM D 7573, and USEPA 415.1.	PM0	No preparation is required.			AR	Yes
TM73	Modified US EPA methods 150.1 and 9045D and BS1377:1990. Determination of pH by Metrohm automated probe analyser.	PM11	Extraction of as received solid samples using one part solid to 2.5 parts deionised water.	Yes	Yes	AR	No
TM77	Modified DDCE/NTS method 15364:2006. Determination of Acid Neutralization Capacity by Metrohm automated probe analyser.	PM0	No preparation is required.			AR	No
TM173	Analysis of fluoride by ISE (Ion Selective Electrode) using modified ISE method 340.2	PM0	No preparation is required.			AR	Yes
NONE	No Method Code	PM4	Gravimetric measurement of Natural Moisture Content and % Moisture Content at either 35°C or 105°C. Calculation based on ISO 11465 and BS1377.			AR	